

# CALIBRATION CERTIFICATE

for  
**Kustom LaserCam 4 Speed Meter**

**Customer:** Police Services of Northern Ireland  
**Serial Number:** LE3255  
**Certificate Number:** 8236  
**Date of Calibration:** 08 August 2024  
**Next Calibration:** 07 August 2025  
**Place of Calibration:** 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

## Calibration Results:

Checksum - CD748BAF  
Oscillator Frequency - 19.999544 MHz  
Self Test Sequence - Pass  
External Controls - Pass  
Aiming - Vertical - Pass  
Horizontal - Pass  
Eye Safety - Pass

## Range $\pm 0.2$ m

Actual	47.64 m	108.69 m	148.86 m
Indicated	47.6 m	108.7 m	148.9 m

## Speed $\pm 1$ MPH

Actual	73 MPH	-23 MPH	-50 MPH	65 MPH	39 MPH
Indicated	73 MPH	-23 MPH	-50 MPH	65 MPH	39 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification  
Range 3 m - 750 m, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK

L Korce

Technical Signatory

# TRUVELO (UK) LTD

TRU-CAL-CER-009

Tel: 020 8847 4400

email: [info@truelouk.com](mailto:info@truelouk.com)