

CALIBRATION CERTIFICATE

for
Kustom LaserCam 4 Speed Meter

Customer: Police Services of Northern Ireland
Serial Number: LE3314
Certificate Number: 8630
Date of Calibration: 10 December 2024
Next Calibration: 09 December 2025
Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CD748BAF
Oscillator Frequency - 19.999736 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
 Horizontal - Pass
Eye Safety - Pass

Range ± 0.2 m

Actual	47.64 m	108.69 m	148.86 m
Indicated	47.6 m	108.7 m	148.9 m

Speed ± 1 MPH

Actual	72 MPH	-55 MPH	-40 MPH	24 MPH	61 MPH
Indicated	72 MPH	-55 MPH	-40 MPH	24 MPH	61 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification
Range 3 m - 750 m, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK


L. Korcz
Technical Signatory

TRUVELO (UK) LTD

TRU-CAL-CER-009

Tel: 020 8847 4400

email: info@truelouk.com