

# CALIBRATION CERTIFICATE

for  
**Kustom LaserCam 4 Speed Meter**

**Customer:** Police Services of Northern  
Ireland

**Serial Number:** LE3315

**Certificate Number:** 8459

**Date of Calibration:** 30 October 2024

**Next Calibration:** 29 October 2025

**Place of Calibration:** 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

## Calibration Results:

Checksum	-	CD748BAF
Oscillator Frequency	-	19.999799 MHz
Self Test Sequence	-	Pass
External Controls	-	Pass
Aiming - Vertical	-	Pass
Aiming - Horizontal	-	Pass
Eye Safety	-	Pass

Range  $\pm 0.2$  m

Actual	47.64 m	108.69 m	148.86 m
Indicated	47.6 m	108.7 m	148.9 m

Speed  $\pm 1$  MPH


Actual	25 MPH	-72 MPH	-59 MPH	43 MPH	33 MPH
Indicated	25 MPH	-72 MPH	-59 MPH	43 MPH	33 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification  
Range 3 m - 750 m, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK

  
L. Kocz  
Technical Signatory

## TRUVELO (UK) LTD

TRU-CAL-CER-009

Tel: 020 8847 4400

email: [info@truvelouk.com](mailto:info@truvelouk.com)