CALIBRATION CERTIFICATE

Kustom LASERwitness Lite

Mobile Speed Enforcement System

Customer:

Police Services of Northern

Serial Number:

LW06306

Ireland

Certificate Number:

8761

Date of Calibration:

06 February 2025

Next Calibration:

05 February 2026

Place of Calibration:

14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum

- CC85

Oscillator Frequency

- 19.999688 MHz

Self Test Sequence

- Pass

External Controls

- Pass

Aiming

- Pass Vertical Horizontal - Pass

Eye Safety

- Pass

Range ± 0.5 ft

Actual

156.3 ft

356.6 ft

488.4 ft

Indicated

156.4 ft

356.8 ft

488.7 ft

Speed ± 1 MPH

Actual

30 MPH

41 MPH

-74 MPH

-59 MPH

22 MPH

Indicated

30 MPH

41 MPH

-74 MPH

-59 MPH

22 MPH

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification Range 10 ft - 4500 ft, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK

Vechnical Signatory

TRUVELO (UK) LTD

TRU-CAL-CER-005

Tel:020 8847 4400

email:info@truvelouk.com