

CALIBRATION CERTIFICATE

for
Kustom LASERwitness Lite
Mobile Speed Enforcement System

Customer: Police Services of Northern Ireland
Serial Number: LW06308
Certificate Number: 9731
Date of Calibration: 17 March 2026
Next Calibration: 16 March 2027
Place of Calibration: 14 Brentford Business Centre, Commerce Road, Brentford, TW8 8LG

Calibration Results:

Checksum - CC85
Oscillator Frequency - 19.999958 MHz
Self Test Sequence - Pass
External Controls - Pass
Aiming - Vertical - Pass
Horizontal - Pass
Eye Safety - Pass

Range ± 0.5 ft

| | | | |
|-----------|----------|----------|----------|
| Actual | 156.3 ft | 356.6 ft | 488.4 ft |
| Indicated | 156.2 ft | 356.7 ft | 488.7 ft |

Speed ± 1 MPH

| | | | | | |
|-----------|---------|---------|--------|---------|---------|
| Actual | -25 MPH | -50 MPH | 75 MPH | 100 MPH | 125 MPH |
| Indicated | -25 MPH | -50 MPH | 75 MPH | 100 MPH | 125 MPH |

The sample measurements in the calibration certify this device for use to the full extent of the instrument specification
Range 10 ft - 4500 ft, Speed 0 mph - 200 mph

The Laser speedmeter detailed above has been calibrated to Home Office requirements, test instruments calibrated to standards traceable to NPL.

We certify that if any parts were replaced during the calibration process that they are identical to those specified for use in the approved device. This device conforms with the device that is approved.

For Truvelo UK


L Korcz
Technical Signatory

TRUVELO (UK) LTD

